

法定计量检定机构
Legal Metrological Verification Institute
无锡市计量测试院

Wuxi Institute of Metrology and Testing

校准证书

Calibration Certificate

证书编号: 21894483
Certificate No.

送检单位
Applicant

无锡允新中测量仪有限公司

联络信息
Contact information

无锡新区高浪东路19号第一上海广场510

计量器具名称
Name of Instrument

螺纹环规

型号 / 规格
Type/Specification

MJB217×1.5

出厂编号
Serial No.

1#

制造单位
Manufacturer

/

接收日期
Date of Accept

2021年10月11日



授权签字人
Approved by

万凤鸣

万凤鸣

部长

核验员
Checked by

董良海

董良海

校准员
Calibrated by

毛晓峰

毛晓峰

校准日期 2021 年 10 月 11 日
Calibration Date Year Month Day

签发日期 2021 年 10 月 11 日
Issued Date Year Month Day

地址: 江苏省无锡市春新东路8号
Add: No.8 Chunxin East Road, Wuxi, Jiangsu
网址 (HTTP): www.wxmtc.com

电话 (Tel): 0510-82803164 传真 (Fax): 0510-82805098

邮编 (Post Code): 214101 电子邮件 (Email): wxmtc@wxmtc.com





本院是国家法定计量检定机构, 计量授权证书号: (苏) 法计(2018)1002号。
This laboratory is national legal metrological verification institute. Authorization certificate No. (2018)1002.

本院依据CNAS-CL01《检测和校准实验室能力的认可准则》(ISO/IEC17025), 建立质量管理体系。
This laboratory has built up quality management system according to CNAS-CL01 "Accreditation Criteria for the Competence of Testing and Calibration Laboratories" (ISO/IEC 17025).

本次校准的技术依据:
Reference documents for the calibration:

参照 JJF 1345-2012《圆柱螺纹量规校准规范》
客户图纸技术要求

本次校准所使用的主要计量器具:
Main measurement standard used in the calibration:

名称 Name	测量范围 Measurement Range	不确定度/准确度等级/ 最大允许误差 Uncertainty/Accuracy Class/ MPE	证书编号/溯源机构/ 有效期 Certificate No./Institute/ Valid Date
数显测长仪	(0~1000) mm	MPE: $\pm (0.15 + L/2000) \mu m$	21842499/无锡市计量测试院/2022-8-25

校准地点和环境条件:
Place and environmental conditions of the calibration:

地点: 无锡市计量测试院111室
Place
温度: 20℃
Temperature
相对湿度: 58%
R. Humidity

- 注: 1、未经本院书面授权, 不得部分复制本证书。
Note: This certificate shall not be reproduced except in full, without the written approval of the laboratory.
- 2、本院对证书中的所有信息负责, 客户提供的信息除外。
This laboratory is responsible for all information in the calibration certificate, except for the information provided by the applicant.
- 3、本证书的校准结果仅对所校准仪器有效。
The results are only responsible for the items calibrated.
- 4、本证书未加盖校准专用章无效。
The certificate will be invalid if it is issued without official stamp.



校准结果

Results of the calibration

项目 Items	技术要求(mm) Technical requirement		中径实测值 (mm) Measured value of pitch diameter	测量不确定度 Uncertainty (mm) U ($k=2$)	结论 Pass/ Fail
	上限值 Upper value	下限值 Lower value			
通端 (GO)	215.932	215.906	215.932	0.006	符合/pass
止端 (NG)	215.774	215.748	215.771	0.006	符合/pass

敬告:

Suggestions

1.被校准仪器修理后,应立即进行校准。

Recalibrate the instrument immediately after it has been repaired.

2.在使用过程中,如对被校准仪器的技术指标产生怀疑,请重新校准。

Recalibrate the instrument when any suspicion about its performance arises.

3.根据客户要求或校准文件的规定,通常情况下12个月校准一次。

According to the client orders or the calibration documents, the calibration will be done every 12 months as a rule.

以下空白/Following blank